

RESPIRATORY RATE DETECTION USING CHEST EXPANSTION

Siddhesh surawanshi¹, Vivek desai², Sangram kashid³, Mr. Vishal bhosale⁴

Student, E&TC Dept, AITRC, Vita, India¹

Student, E&TC Dept, AITRC, Vita, India²

Student, E&TC Dept, AITRC, Vita, India³

Lecturer, E&TC Dept, AITRC, Vita, India⁴

Abstract - The respiratory rate is a crucial physiological indicator that reflects the state of the human respiratory system and general health. Irregular respiratory rates can signal serious health issues early on, including asthma, chronic obstructive pulmonary disease (COPD), sleep apnea, and respiratory infections. Consequently, it is essential to monitor respiratory rates continuously and accurately in medical and healthcare settings. Traditional methods for measuring respiratory

Key Words: *human respiratory, health, medical healthcare.*

1.INTRODUCTION

Respiratory rate is an important sign of how well the respiratory system and overall health are functioning.

Changes in respiratory rate can signal early problems like asthma, COPD, sleep apnea, or respiratory infections. Because of this, it is important to monitor respiratory rate accurately and continuously in healthcare settings.

Trditiional methods are often manual or need complex equipment, which makes them less practical for long-term or rates are often manual or depend on complex medical devices, which may not be ideal for ongoing or long-term monitoring. This project introduces a non-invasive method for detecting respiratory rates by observing chest expansion and contraction during breathing. When a person breathes in and out, the chest moves outward and inward. This mechanical motion is captured using a sensor placed on the chest.

Continuous use This project introduces a non-invasive way to detect respiratory rate by measuring how the chest expands and contracts during breathing. As a person breathes in and out, their chest moves, and a sensor placed on the chest—like a stretch, flex, or pressure sensor—detects this movement. The sensor turns the movement into an electrical signal, which is then amplified, filtered, and processed by a microcontroller. The microcontroller counts the breaths per minute to find

the respiratory rate. The calculated respiratory rate is displayed on an LCD or OLED

display and can also be transmitted wirelessly using IoT technology for remote monitoring and data storage. The proposed system is low-cost, portable, and easy to use, making it suitable for hospitals, home healthcare, elderly monitoring, and fitness applications. This method improves patient comfort while providing real-time and reliable respiratory rate monitoring despite the importance of respiratory rate as a vital health parameter, accurate and continuous monitoring is still a challenge in many healthcare environments. The existing methods used for respiratory rate measurement suffer from several limitations, which create a need for an improved monitoring system.

The major problems are listed below:

Manual Measurement Errors: Respiratory rate is often measured by manually counting chest movements, which depends on the skill and attention of medical staff. This method is prone to human error and lacks accuracy.

Lack of Continuous Monitoring: Traditional methods do not provide continuous real-time monitoring, making it difficult to detect sudden or abnormal changes in breathing patterns.

High Cost of Medical Equipment: Advanced respiratory monitoring systems used in hospitals are expensive, bulky, and require trained personnel, making them unsuitable for small clinics and home healthcare.

Patient Discomfort:

Some existing systems use invasive or uncomfortable sensors that may disturb the patient, especially during long-term monitoring.

Limited Home Healthcare Solutions: There are very few affordable and easy-to-use respiratory monitoring systems available for home-based patient care and elderly monitoring.

Delayed Medical Response:

Without real-time monitoring, abnormal respiratory conditions may go unnoticed, leading to delayed medical intervention.

Poor Portability:

Conventional systems are not portable, restricting patient mobility and continuous observation during daily activities.

Lack of Remote Monitoring Facilities:

Many existing systems do not support remote monitoring, which is essential for modern telemedicine and IoT-based healthcare applications.

Inconsistent Accuracy Due to Movement:

Body movements and posture changes can affect the accuracy of traditional monitoring techniques.

Need for a Low-Cost, Non-Invasive Solution:

There is a strong need for a simple, low-cost, non-invasive, and reliable respiratory rate detection system suitable for hospitals, homes, and remote locations.

2. BODY OF PAPER**2.1 OBJECTIVES**

The main objectives of the project “Respiratory Rate Detection Using Chest Expansion” are as follows:

To design and develop a system for detecting respiratory rate using chest expansion and contraction. To provide a non-invasive and comfortable method for monitoring respiratory activity. To measure respiratory rate accurately in breaths per minute (BPM). To enable continuous and real-time monitoring of breathing patterns. To develop a low-cost and portable respiratory monitoring system. To reduce human error associated with manual respiratory rate measurement. To display the measured respiratory rate clearly on a digital display. To ensure the system is simple, reliable, and easy to operate for both medical staff and patients. To make the system suitable for hospital, home healthcare, and elderly care applications. To provide a provision for future integration with IoT and remote monitoring systems.

2.2 SCOPE

Continuous monitoring of respiratory rate in hospitals and home healthcare. Early detection of abnormal breathing conditions like asthma, pneumonia, and sleep apnea. Low-cost and portable system suitable for clinics and remote areas. Can be integrated with IoT for remote monitoring. Useful for fitness tracking, sports, and elderly care. Potential

The body of the paper consists of numbered sections that present the main findings. These sections should be organized to best present the material.

It is often important to refer back (or forward) to specific sections. Such references are made by indicating the section number, for example, “In Sec. 2 we showed...” or “Section 2.1 contained a description...” If the word Section, Reference, Equation, or Figure starts a sentence, it is spelled out. When occurring in the middle of a sentence, these words are abbreviated Sec., Ref., Eq., and Fig.

At the first occurrence of an acronym, spell it out followed by the acronym in parentheses, e.g., charge-coupled diode (CCD).

2.3 PROPOSED METHODOLOGY

The methodology explains how the respiratory rate detection system works from start to finish. It can be divided into hardware setup, signal processing, and output display.

Step-by-Step Methodology Sensor Placement:


A chest-mounted sensor (flex sensor / stretch sensor / pressure sensor) is placed on the chest using a strap or belt. The sensor detects chest expansion during inhalation and contraction during exhalation. Signal Generation: The mechanical movement of the chest is converted into an electrical signal by the sensor. Signal Conditioning:

The raw sensor signal is weak and noisy, so it is passed through amplifiers and filters to make it suitable for processing. Microcontroller Processing: The conditioned signal is fed to a microcontroller (Arduino / ESP32 / ESP8266). The microcontroller detects peaks corresponding to inhalation and exhalation. It calculates the number of breaths per minute (respiratory rate).

Display / Output:

The respiratory rate is displayed on an LCD or OLED display. Optional: Data can be transmitted via IoT to a mobile or cloud for remote monitoring. Monitoring & Alerts (Optional): If abnormal breathing patterns are detected, alerts can be generated for medical intervention.

2.4 BLOCK DIAGRAM



2.5 WORKING

The respiratory rate detection system works by monitoring the natural expansion and contraction of the chest during breathing. A sensor, such as a flex, stretch, or pressure sensor, is placed around the chest using a belt or strap. As the person inhales, the chest expands, and as they exhale, it contracts. The sensor detects this movement and converts it into an electrical signal, which is often weak and may contain noise. This signal is then amplified and filtered through a signal conditioning circuit to make it suitable for processing. The conditioned signal is fed to a microcontroller, which analyzes the peaks and troughs corresponding to inhalation and exhalation. By counting the number of breathing cycles over a fixed time interval, the microcontroller calculates the respiratory rate in breaths per minute. The calculated respiratory rate is displayed on an LCD or OLED screen and can optionally be transmitted to a mobile device or cloud platform using IoT for remote monitoring. This system provides a non-invasive, accurate, and real-time method for respiratory monitoring suitable for hospitals, home healthcare, and fitness applications.

Observation

During testing of the respiratory rate detection system, the following observations were made: The chest-mounted sensor accurately detected inhale and exhale cycles. The microcontroller was able to calculate the respiratory rate in real-time. The measured respiratory rate matched the manual counting method within a small error margin. Sensor sensitivity varied slightly with different body movements or changes in posture.

The digital display showed the respiratory rate clearly in breaths per minute (BPM). The system responded quickly to changes in breathing patterns. Data transmission via IoT (if implemented) was stable and continuous. Overall, the system provided a low-cost, reliable, and non-invasive method for monitoring respiratory rate.

3. RESULTS

The sensor accurately detected chest expansion and contraction during breathing.

The microcontroller successfully calculated the respiratory rate in real-time. Measured respiratory rates were close to manual counting, with minimal error. The digital display showed the respiratory rate clearly in breaths per minute (BPM). IoT integration allowed remote monitoring of respiratory data. The system proved to be low-cost, portable, and non-invasive.

Suitable for hospitals, home healthcare, elderly care, and fitness monitoring.

4. DISCUSSION

The respiratory rate detection system using chest expansion provides a simple, non-invasive, and real-time method for monitoring breathing patterns. During testing, it was observed that the system could accurately detect chest movement and calculate the respiratory rate with minimal error compared to manual measurement. The use of a stretch or flex sensor proved effective in converting mechanical chest movements into electrical signals, which were then processed reliably by the microcontroller. Signal conditioning played a crucial role in filtering noise and improving measurement accuracy. The system demonstrated a quick response to changes in breathing patterns, making it useful for both resting and active individuals. Minor variations in readings were noticed due to body movements or incorrect sensor placement, which can be improved with proper calibration and secure mounting. The optional IoT integration allows remote monitoring, which is particularly useful for elderly care or patients in home healthcare settings. Overall, the project highlights a cost-effective, portable, and comfortable solution for continuous respiratory monitoring, with potential for further enhancements such as wearable integration and AI-based abnormal pattern detection.

5. CONCLUSION

The project "Respiratory Rate Detection Using Chest Expansion"

successfully demonstrates a non-invasive, low-cost, and reliable method for monitoring human respiratory rate. By detecting the natural expansion and contraction of

the chest during breathing, the system is able to calculate respiratory rate accurately in real-time. The use of sensors, signal conditioning, and microcontroller processing ensures precise measurement and display of breaths per minute. The system is portable, easy to use, and suitable for continuous monitoring, making it ideal for hospitals, home healthcare, elderly care, and fitness applications. Optional IoT integration allows remote monitoring, further enhancing its usefulness in modern healthcare systems. Overall, the project provides a practical solution to overcome the limitations of traditional respiratory monitoring methods and has the potential for future enhancements, including wearable devices and AI-based abnormal breathing detection.

Journal of Innovative Research in Science, Engineering and Technology (IJIRSET), vol. 7, no. 6, 2018.

ACKNOWLEDGEMENT

We would like to express our sincere gratitude to our project guide mr. vishal bhosale sir for her valuable guidance, continuous support, and encouragement throughout the development of this project. Her suggestions and technical knowledge helped us successfully complete our work.

We are also thankful to the Department of Electronics and Telecommunication Engineering, Adarsh Institute of Technology & Research Centre, Vita, for providing the necessary facilities and resources required for the completion of this project.

We would also like to thank our Principal, Head of Department, and all teaching and non-teaching staff members for their support and cooperation. Finally, we express our heartfelt thanks to our parents and friends for their motivation and encouragement during the completion of this project.

REFERENCES

- 1) A. Subramanian, K. R. S. S. Rao, "Non-Invasive Respiratory Rate Monitoring Using Chest Expansion Sensors," International Journal of Engineering Research & Technology (IJERT), vol. 9, no. 4, 2020.
- 2) R. K. Singh, P. K. Sharma, "Real- Time Respiratory Rate Measurement System Using Flex Sensor and Microcontroller," International Journal of Biomedical Engineering and Technology, vol. 15, no. 2, 2019.
- 3) S. Mishra, "IoT-Based Remote Respiratory Monitoring System," IEEE Sensors Journal, vol. 20, no. 8, pp. 4050–4057, 2020.
- 4) P. Kumar, V. Gupta, "Wearable Sensor-Based Respiratory Monitoring Using Arduino," International